

C E R T I F I C A T E   O F   C O M P L I A N C E

LOT #: 639901

KEY SCIENTIFIC PRODUCTS  
1113 E REYNOLDS STREET  
STAMFORD, TEXAS 79553  
1-800-THE-1-KEY

THIS LOT HAS BEEN TESTED ACCORDING TO THE CRITERIA LISTED  
BELOW AND MEETS THE PERFORMANCE STANDARDS AS ESTABLISHED  
BY KEY SCIENTIFIC PRODUCTS TO MEET NCCLS GUIDELINES.

PRODUCT: HIPPURATE DISCS 50/BTL  
CATALOG #: K710310

LOT #: 639901

PACKAGED DATE = 2023/12/08    EXP DATE = 2025/12/08

CONTROL LOT# 580800

TESTED BY: BLC

		POSITIVE CURRENT	NEGATIVE CURRENT	POSITIVE CONTROL	NEGATIVE CONTROL
12386	GROUP B STREP	+		+	
19615	STREPTOCOCCUS PYOGEN		-		-

REAGENT # 614100

DISK COLOR WHITE    INCUBATE FOR 2 HOURS

POSITIVE: PURPLE COLOR AFTER ADDITION OF DEVELOPER.

NEGATIVE: NO COLOR CHANGE AFTER ADDITION OF DEVELOPER.

CHECK HIPPURATE VOL /DISK

WEIGH 10 BLANK DISKS OF SAME SIZE

WEIGH 10 FINISHED HIPPURATE DISKS

DIFFERENCE SHD BE 1.8 TO 2.2 MG PER DISK