

C E R T I F I C A T E O F C O M P L I A N C E

LOT #: 615201

KEY SCIENTIFIC PRODUCTS
1113 E REYNOLDS STREET
STAMFORD, TEXAS 79553
1-800-THE-1-KEY

THIS LOT HAS BEEN TESTED ACCORDING TO THE CRITERIA LISTED
BELOW AND MEETS THE PERFORMANCE STANDARDS AS ESTABLISHED
BY KEY SCIENTIFIC PRODUCTS TO MEET NCCLS GUIDELINES.

PRODUCT: PYR DISCS (W/PEP REAGENT)
CATALOG #: K1538B

LOT #: 615201

PACKAGED DATE = 2022/05/24 EXP DATE = 2024/05/20

CONTROL LOT# 610600

TESTED BY: BLC

REAGENT USED-PEP: 6023,6062

	CONTROL	THIS LOT
13813	STREPTOCOCCUS AGALAC	NEG
29212	ENTEROCOCCUS FAECALI	POS
19615	STREPTOCOCCUS PYOGEN	POS
25923	STAPHYLOCOCCUS AUREU	NEG
13882	KLEBSIELLA PNEUMONIA	POS
25922	ESCHERICHIA COLI	NEG
13048	ENTEROBACTER AEROGEN	POS
8100	SERRATIA MARCESSENS	POS

CONTROL STRIPS 1-24
ALL STRIPS PASSED

POSITIVE TEST IS RED OR DARK PINK AFTER ADDING
REAGENT. GREEN, YELLOW, OR VERY PALE PINK IS NEGATIVE.
TESTED CONTROL STRIP FROM EACH TUB AND RANDOM DISCS.