

C E R T I F I C A T E   O F   C O M P L I A N C E

LOT #: 615203

KEY SCIENTIFIC PRODUCTS  
1113 E REYNOLDS STREET  
STAMFORD, TEXAS 79553  
1-800-THE-1-KEY

THIS LOT HAS BEEN TESTED ACCORDING TO THE CRITERIA LISTED  
BELOW AND MEETS THE PERFORMANCE STANDARDS AS ESTABLISHED  
BY KEY SCIENTIFIC PRODUCTS TO MEET NCCLS GUIDELINES.

PRODUCT: PYR DISCS (W/PEP REAGENT)  
CATALOG #: K1538B

LOT #: 615203

PACKAGED DATE = 2022/08/22    EXP DATE = 2024/08/17

CONTROL LOT# 610600

TESTED BY: BLC

REAGENT USED-PEP:    6023,6062

	CONTROL	THIS LOT	
13813	STREPTOCOCCUS AGALAC	NEG	NEG
29212	ENTEROCOCCUS FAECALI	POS	POS
19615	STREPTOCOCCUS PYOGEN	POS	POS
25923	STAPHYLOCOCCUS AUREU	NEG	NEG
13882	KLEBSIELLA PNEUMONIA	POS	POS
25922	ESCHERICHIA COLI	NEG	NEG
13048	ENTEROBACTER AEROGEN	POS	POS
8100	SERRATIA MARCESSENS	POS	POS

CONTROL STRIPS 1-24  
ALL STRIPS PASSED

POSITIVE TEST IS RED OR DARK PINK AFTER ADDING  
REAGENT. GREEN, YELLOW, OR VERY PALE PINK IS NEGATIVE.  
TESTED CONTROL STRIP FROM EACH TUB AND RANDOM DISCS.