

C E R T I F I C A T E O F C O M P L I A N C E

LOT #: 661304

KEY SCIENTIFIC PRODUCTS
1113 E REYNOLDS STREET
STAMFORD, TEXAS 79553
1-800-THE-1-KEY

THIS LOT HAS BEEN TESTED ACCORDING TO THE CRITERIA LISTED
BELOW AND MEETS THE PERFORMANCE STANDARDS AS ESTABLISHED
BY KEY SCIENTIFIC PRODUCTS TO MEET NCCLS GUIDELINES.

PRODUCT: HIPPURATE DISCS 50/BTL
CATALOG #: K710310

LOT #: 661304

PACKAGED DATE = 2026/01/23 EXP DATE = 2028/01/23

CONTROL LOT# 639900

TESTED BY: BLC

HIPP DEV LOT 633400

		POSITIVE CURRENT	NEGATIVE CURRENT	POSITIVE CONTROL	NEGATIVE CONTROL
12386	GROUP B STREP	+	-	+	-
19615	STREPTOCOCCUS PYOGEN				

DISK COLOR WHITE INCUBATE FOR 2 HOURS
POSITIVE: PURPLE COLOR AFTER ADDITION OF DEVELOPER.
NEGATIVE: NO COLOR CHANGE AFTER ADDITION OF DEVELOPER.
CHECK HIPPURATE VOL /DISK
WEIGH 10 BLANK DISKS OF SAME SIZE
WEIGH 10 FINISHED HIPPURATE DISKS
DIFFERENCE SHD BE 1.8 TO 2.2 MG PER DISK